

الشروط الخاصة التسلسل

- 1 - Goods including injectors must be freshly prepared on dispatch/delivery and must retain at least 2/3 of its shelf life when dispatched /delivered. The DRMS with this respect reserves the right to reject any item not in compliance with this term and impose the correspondent fines as per GHQ /Military procurement laws .
- 2 - Dispatched/delivered goods must show description of goods, manufacturing date, expiry date, lot number, storage conditions and any other necessary information. Where the manufacturing date is not applicable, it should be provided on a separate document.
- 3 - Goods must be dispatch under the same storage conditions that comply with their nature & Storage Conditions Must be Clearly Indicated on Packages as Well as in the Relevant Documents.
- 4 - Quoted lenses should be FDA or CE approved.
- 5 - Quoted lenses should be previously approved by DRMS, and supportive documents should be provided with the offers.
- 6 - DRMS has the right to exchange lenses damaged during surgical operations if it is due to manufacturing default.
- 7 - DRMS has the right to exchange any power of the awarded lenses with any other power through the local agent, such exchange should be affected within a week after receipt of our written request. Any delay will impose penalties as per GHQ/RMS regulations.

بسم الله الرحمن الرحيم

القيادة العامة للقوات المسلحة الاردنية

مديرية الخدمات الطبية الملكية

التزويد

صفحة رقم 2 - 3

الشروط الخاصة للعطاء رقم P17-2017-B-3 العدسات المطوية

الشروط الخاصة التسلسل

- | | |
|----|--|
| 8 | - After arrival to Jordan, DRMS is not responsible for any handling charges for goods (like charges for issuing delivery note vouchers, unloading charges, local shipping agents charges). Furthermore, DRMS is not responsible for any demurrage charges and other charges incurred by the port corporation caused by delay in presenting all necessary shipping documents for clearing of goods. |
| 9 | - Shipment/ delivery: in Two Shipments/deliveries (4-5) months apart. |
| 10 | - All parcels and packages and related documents should be marked: DRMS, tender no. P17-2015-B-7. GHQ/ Jordan Armed Forces, Amman, Jordan. |
| 11 | - DRMS has the right to increase the awarded quantities by a percentage not exceeding 30% after final order with the same prices, terms and conditions. |
| 12 | - The local agent should has all powers from zero to 30 dioptr with increasing power 0.5 |
| 13 | - Prices must include services of sale, shipment, transportation, delivery from port to site or Main Medical Stores. |

التسلسل الشروط الخاصة

- 14 - Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), how ever, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation or QAIA (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges) .The supplier is also responsible for providing af all relevant shipping documents ,together with the delivery orders.
- 15 - SAMPLES AND CATALOGUES OF THE QUOTED ITEM MUST BE SUBMITTED ALONG WITH THE OFFERS FOR PURCHASING AND RECEIPT PURPOSES SHOWING THE SAME SERIAL NUMBER/ CAT NUMBER AS PER OUR TENDER DOCUMENTS SPECIFICATIONS .RMS WITH THIS RESERVES THE RIGHT TO REJECT OFFERS NOT IN COMPLIANCE WITH THIS TERM.
- 16 - For items no.(41-68) bidders are requested to provide offers of both hydrophilic and hydrophobic three piece lenses and DRMS has the right to purchase according to our needs
- 17 - For items no.(1-40) injectors (Qty:150 each) should be submitted on F.O.C basis.
- 18 - In case of the required dioptr, A-constant and optic size asked to be exchanged is not provided, DRMS with this regard has the right to buy the suitable lenses from any other sources at your expenses and will incur the correspondent fine as per GHQ/ military procurement law.

DIRECTORATE ROYAL MEDICAL SERVICES

العدسات المطوية

TENDER NO P17-2017-B-3

Foldable Intraocular Lenses:-
 1-I.O.L Material:Acrylic.
 2-U.V Absorbable.
 3-One piece Lenses.
 4-Optic Diameter: 5.75mm-6.0mm.
 5-Optic Configuration: Biconvex.
 6-A/Costant 116-119.5.
 7-Haptics:0 Angle.
 8-I.O.L with square edge optic
 9.Injectable including cartillage
 and injector in 3.0mm or less incision
 10-aspheric design.

1	20_14_003_100	12 EACH	Power: 10.0 Diopters.
2	20_14_003_105	2 EACH	Power: 10.5 Diopters.
3	20_14_003_110	12 EACH	Power: 11.0 Diopters.
4	20_14_003_115	4 EACH	Power: 11.5 Diopters.
5	20_14_003_120	12 EACH	Power: 12.0 Diopters.
6	20_14_003_125	2 EACH	Power: 12.5 Diopters.
7	20_14_003_130	12 EACH	Power: 13.0 Diopters.
8	20_14_003_135	4 EACH	Power: 13.5 Diopters.
9	20_14_003_140	12 EACH	Power: 14.0 Diopters.
10	20_14_003_145	6 EACH	Power: 14.5 Diopters.
11	20_14_003_150	18 EACH	Power: 15.0 Diopters.
12	20_14_003_155	4 EACH	Power: 15.5 Diopters.
13	20_14_003_160	20 EACH	Power: 16.0 Diopters.
14	20_14_003_165	12 EACH	Power: 16.5 Diopters.
15	20_14_003_170	40 EACH	Power: 17.0 Diopters.
16	20_14_003_175	12 EACH	Power: 17.5 Diopters.

DIRECTORATE ROYAL MEDICAL SERVICES

العدسات المطوية

TENDER NO P17-2017-B-3

17	20_14_003_180	192 EACH	Power: 18.0 Diopters.
18	20_14_003_185	190 EACH	Power: 18.5 Diopters.
19	20_14_003_190	384 EACH	Power: 19.0 Diopters.
20	20_14_003_195	262 EACH	Power: 19.5 Diopters.
21	20_14_003_200	626 EACH	Power: 20.0 Diopters.
22	20_14_003_205	562 EACH	Power: 20.5 Diopters.
23	20_14_003_210	628 EACH	Power: 21.0 Diopters.
24	20_14_003_215	564 EACH	Power: 21.5 Diopters.
25	20_14_003_220	628 EACH	Power: 22.0 Diopters.
26	20_14_003_225	562 EACH	Power: 22.5 Diopters.
27	20_14_003_230	504 EACH	Power: 23.0 Diopters.
28	20_14_003_235	370 EACH	Power: 23.5 Diopters.
29	20_14_003_240	370 EACH	Power: 24.0 Diopters.
30	20_14_003_245	180 EACH	Power: 24.5 Diopters.
31	20_14_003_250	108 EACH	Power: 25.0 Diopters.
32	20_14_003_255	30 EACH	Power: 25.5 Diopters.
33	20_14_003_260	50 EACH	Power: 26.0 Diopters.
34	20_14_003_265	14 EACH	Power: 26.5 Diopters.
35	20_14_003_270	10 EACH	Power: 27.0 Diopters.
36	20_14_003_275	4 EACH	Power: 27.5 Diopters.
37	20_14_003_280	6 EACH	Power: 28.0 Diopters.

Sunday, May 21 2017

Page 3 of 5

DIRECTORATE ROYAL MEDICAL SERVICES

العدسات المطوية

TENDER NO P17-2017-B-3

[REDACTED]			
38	20_14_003_285	4 EACH	Power: 28.5 Diopters.
39	20_14_003_290	6 EACH	Power: 29.0 Diopters.
40	20_14_003_300	6 EACH	Power: 30.0 Diopters.

DIRECTORATE ROYAL MEDICAL SERVICES

العدسات المطوية

TENDER NO P17-2017-B-3

Foldable Intraocular Lenses
 Intraocular lenses 1-I.O.L Material:Hydrophobic or
 Hydrophilic,Acrylic with U.V and blue light filtering
 chromophore (Methylphenylazo hydroxyphenyl
 Ethylmethacrylamide):
 1.Three pieces with stableforce haptic
 2.Optic diameter 6.0 mm
 3.Over length 13.0mm
 4.Refractive index: 1.44-1.55
 5.Aspheric design
 6.A/constant :116.5-119.5
 7.Injectable including cartilage and injector in 3.0mm
 or less incision

41	20_14_00313_16 0	20 EACH	Diopter diameter 16.0
42	20_14_00313_16 5	10 EACH	Diopter diameter 16.5
43	20_14_00313_17 0	20 EACH	Diopter diameter 17.0
44	20_14_00313_17 5	10 EACH	Diopter diameter 17.5
45	20_14_00313_18 0	24 EACH	Diopter diameter 18.0
46	20_14_00313_18 5	36 EACH	Diopter diameter 18.5
47	20_14_00313_19 0	38 EACH	Diopter diameter 19.0
48	20_14_00313_19 5	40 EACH	Diopter diameter 19.5
49	20_14_00313_20 0	52 EACH	Diopter diameter 20.0
50	20_14_00313_20 5	52 EACH	Diopter diameter 20.5
51	20_14_00313_21 0	52 EACH	Diopter diameter 21.0
52	20_14_00313_21 5	52 EACH	Diopter diameter 21.5
53	20_14_00313_22 0	52 EACH	Diopter diameter 22.0
54	20_14_00313_22 5	40 EACH	Diopter diameter 22.5
55	20_14_00313_23 0	40 EACH	Diopter diameter 23.0
56	20_14_00313_23 5	24 EACH	Diopter diameter 23.5
57	20_14_00313_24	24 EACH	Diopter diameter 24.0

DIRECTORATE ROYAL MEDICAL SERVICES

العدسات المطوية

TENDER NO P17-2017-B-3

58	20_14_00313_24 5	12 EACH	Dioptr diameter 24.5
59	20_14_00313_25 0	24 EACH	Dioptr diameter 25.0
60	20_14_00313_25 5	12 EACH	Dioptr diameter 25.5
61	20_14_00313_26 0	20 EACH	Dioptr diameter 26.0
62	20_14_00313_26 5	10 EACH	Dioptr diameter 26.5
63	20_14_00313_27 0	10 EACH	Dioptr diameter 27.0
64	20_14_00313_27 5	10 EACH	Dioptr diameter 27.5
65	20_14_00313_28 0	10 EACH	Dioptr diameter 28.0
66	20_14_00313_28 5	10 EACH	Dioptr diameter 28.5
67	20_14_00313_29 0	10 EACH	Dioptr diameter 29.0
68	20_14_00313_30 0	6 EACH	Dioptr diameter 30.0