

## DIRECTORATE ROYAL MEDICAL SERVICES

مستهلكات الكلية الإصطناعية

TENDER NO P7-2017

**16\_03 DIALYSIS DISPOSABLES**

			Haemo Dialysis Bicarbonate Solution 8.4%
9	16_03_040	133000 GALL	5-6 Lit/Gallon.
			Haemo Dialysis Concentrated Solution
10	16_03_042	264000 GALL	Contents Per Liter :Nacl 263 gm, Ca Chloride 2H <sub>2</sub> O 9.9 gm, Mg Chloride. 6H <sub>2</sub> O 3.4 gm. KCL 3.4gm , Acetic Acid 8.1 gm. Dextrose Anhydrous 90 gm. For Bicarbonate Dialysis (3.78 Lt/Gallon ).
			Hollow Fiber Dialyzer fully synthetic, low flux
11	16_03_0511	59000 EACH	surface area (1.5-1.7 m <sup>2</sup> )
12	16_03_0512	160000 EACH	surface area (1.8-1.9 m <sup>2</sup> )
13	16_03_0513	99000 EACH	surface area (2-2.1 m <sup>2</sup> )
			Hollow Fiber Dialyzer fully synthetic surface area (0.6m <sup>2</sup> )
14	16_03_052	4400 EACH	LOW FLUX
			Hollow Fiber Dialyzer fully synthetic surface area (0.9-1m <sup>2</sup> )
15	16_03_0522	9000 EACH	High flux
			Adjustable Spring-Loaded Kidney Biopsy Needle,
16	16_03_0591	70 EACH	G (16-20) Length 9
			Needle, Fistula, Arterial- Venous With Backeye, Clamp, Rotatable Needle, length 20-30 cm.
17	16_03_060	125000 EACH	Size G 16
18	16_03_062	100000 EACH	Size G 17
			Equilibration Chamber Filter to fit Fresenius
19	16_03_072	10 EACH	
			Equilibration Chamber Filter to fit Nikkiso
20	16_03_073	170 EACH	

**DIRECTORATE ROYAL MEDICAL SERVICES**

مستهلكات الكلية الإصطناعية

TENDER NO P7-2017

**16\_03 DIALYSIS DISPOSABLES**

		Arterial-Venous bloodline tubing, Adult (To Fit Nikkiso DBB 06).	
4	16_03_01	64000 SET	<p>*Arterial line to include (but not limited to):</p> <ol style="list-style-type: none"> <li>1-blood injection site.</li> <li>2-on off clamp for all connectors (arterial, dialyzer inlet pressure, heparin, dialyzer and patient line).</li> <li>3-self ejecting patient connector.</li> <li>4-heparin line and connector</li> <li>5-chamber for arterial pressure measurement.</li> <li>6-chamber for dialyzer inlet pressure measurement.</li> <li>7-transducer protector (fluid barrier) for arterial pressure measurement connector.</li> <li>8-transducer protector (fluid barrier) for dialyzer inlet pressure measurement connector.</li> <li>9-blood bump segment.</li> <li>10-vented cap for dialyzer connector.</li> <li>11-dialyzer connector.</li> <li>12-recirculating connector.</li> </ol> <p>*Venous line to include (but not limited to):</p> <ol style="list-style-type: none"> <li>1-dialyzer connector.</li> <li>2-blood injection site.</li> <li>3-transducer protector (fluid barrier) for venous pressure measurement connector.</li> <li>4-chamber for venous pressure measurement.</li> <li>5-venous chamber internal filter.</li> <li>6-on off clamp for all connectors (venous, dialyzer and patient line).</li> <li>7-drainage bag.</li> </ol>

Catheter, Subclavian - Femoral Double Lumen Kit 10fr . with curved extentions. Kit includes: introducer needle, guidewire, catheter, injection sealing caps, 10f dilator.

5	16_03_026	2000 EACH	Size 11.5Fr x 13.5cm ( Small Adult ) .
6	16_03_027	2000 EACH	Size 10Fr x 15cm (Medium Adult)
7	16_03_028	400 EACH	Size 10Fr x 12cm (Pediatric)
8	16_03_029	1800 EACH	Size 11.5Fr x 16cm (Large Adult) Form #FP 05-01 Rev. a

## DIRECTORATE ROYAL MEDICAL SERVICES

مستهلكات الكلية الإصطناعية

TENDER NO P7-2017

Item NO	STOCK NO	REQ QTY	UNIT	DESCRIPTION
<b>16_03 DIALYSIS DISPOSABLES</b>				
				Arterial-Venous Line Set For Open System Dialysis Machine With Filter And Clamps, Adult (To Fit Fresenius).
2	16_03_009	38000	SET	A-V blood Tubing set including following accessories post pump arterial chamber .Clamps.Heparin Line,Transducer protection both arterial and venous line 19, Recirculation connector injection post, sigment Pump 6.5 x 9.5.
				A-V blood tubing set including following accessories :Post pump arterial chamber ,Clamps,Heparin Line,Transducer protection both arterial and venous line, Recirculation connector injection post, segment Pump .
3	16_03_0091	220000	SET	Arterial-Venous Line Set For Open System Dialysis Machine With Filter And Clamps, Adult (To Fit B.Braun).

## DIRECTORATE ROYAL MEDICAL SERVICES

مستهلكات الكلية الإصطناعية

TENDER NO P7-2017

Item NO	STOCK NO	REQ QTY	UNIT	DESCRIPTION
16 03				Arterial-Venous Line Set For Open System Dialysis Machine With Filter And Clamps, Pediatric (To Fit B.Braun).
1	16_03_0071	16000	SET	<p>A-V blood tubing set including following accessories :Post pump arterial chamber ,Clamps,Heparin Line,Transducer protection both arterial and venous line, Recirculation connector injection post, segment Pump .</p> <p>From the venous line side: sensor gauge should be adult size to fit the air detector (to fit B. Braun)</p> <p>Clamps to include:</p> <p>I. Venous side: (3) clamps in the following sites:</p> <p>A. Inlet</p> <p>B. Sensor</p> <p>C. Air detector</p> <p>II. Arterial side: (5) clamps in the following sites:</p> <p>A. Outlet</p> <p>B. Pressure sensor</p> <p>C. Saline intake</p> <p>D. PBE sensor</p> <p>E. Heparin pump</p>

بِسْمِ اللَّهِ الرَّحْمَنِ الرَّحِيمِ

القيادة العامة للقوات المسلحة الأردنية

مديرية الخدمات الطبية الملكية

التزويد

صفحة رقم 1 - 3

P7-2017 مستهلكات الكلية الإصطناعية

الشروط الخاصة للعطاء رقم

الشروط الخاصة التسلسل

1	- Goods must be dispatch under the same storage conditions that comply with their nature & Storage Conditions Must be Clearly Indicated on Packages as Well as in the Relevant Documents.
2	- All Outer and Inner Packages To be Printed, Stamped, Or Labeled With The Following "SOLD TO RMS"/ TENDER NO.# P7-2017
3	- The delivered goods must be labelled with the following on the inner & outer packages :description of goods , manufacturing date , expiry date in addition to the batch no , storage conditions and any other necessary information
4	- After arrival to Jordan , DRMS IS not responsible for any handling charges for goods ( like charges for issuing delivery note vouchers unloading charges, local shipping agent's charges).
5	- For items no . (9.10 ),Individual packs should be unbreakable , leak proof containers and outer cartons should be strong , export quality material , able to withstand rough handling during transport and to endure piling up to 10 cartons upon storage , on wooden bases .
6	- For items (9.10) shipments to be scheduled in coordinate with directorate of supply.
7	- All the laboratory investigations for items(9.10) will be at the bidders expenses .

بسم الله الرحمن الرحيم

القيادة العامة للقوات المسلحة الأردنية

مديرية الخدمات الطبية الملكية

التزويد

صفحة رقم 2 - 3

P7-2017 مستهلكات الكلية الإصطناعية

الشروط الخاصة للعطاء رقم

الشروط الخاصة التسلسل

- 8 - Goods must be freshly prepared on dispatch/delivery and must retain at least 2/3 of its shelf life when dispatched /delivered. The DRMS with this respect reserves the right to reject any item not in compliance with this term and impose the correspondent fines as per GHQ /Military procurement laws .
- 9 - All parcels, packages & related documents should be marked:-  
DRMS,TENDER NO.P7/2017 GHQ,JORDAN ARMED FORCES  
AMMAN , JORDAN
- 10 - Goods Should be Previously Approved & Evaluated in DRMS. Approval Certificates (For Previously Approved Items) Should be Submitted along With Offers.
- 11 - SAMPLES AND CATALOGUES OF THE QUOTED ITEM MUST BE SUBMITTED ALONG WITH THE OFFERS FOR PURCHASING AND RECEIPT PURPOSES SHOWING THE SAME SERIAL NUMBER/ CAT NUMBER AS PER OUR TENDER DOCUMENTS SPECIFICATIONS .RMS WITH THIS RESERVES THE RIGHT TO REJECT OFFERS NOT IN COMPLIANCE WITH THIS TERM.

بِسْمِ اللَّهِ الرَّحْمَنِ الرَّحِيمِ

القيادة العامة للقوات المسلحة الأردنية

مديرية الخدمات الطبية الملكية

التزويد

صفحة رقم 3 - 3

P7-2017 مستهلكات الكلية الإصطناعية

الشروط الخاصة للعطاء رقم

الشروط الخاصة التسلسل

12 – Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), how ever, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation or QAIA (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges ) .The supplier is also responsible for providing af all relevant shipping documents ,together with the delivery orders.

13 – Pricing must include services of sale, shipment, transportation, delivery from port to site or to Main Medical Stores.

14 – All items to be delivered in two equal Consignments (8-9) months apart. First Consignment to be within three months.