

DIRECTORATE ROYAL MEDICAL SERVICES

Probe Covers To Fit Digital Thermometer

TENDER NO P45-2018-2

Cover To Fit Digital Thermometer Probe For Rectal
Use

Cover To Fit Digital Thermometer Probe For Rectal Use

1 11_0302_061 200000 EACH

Cover To Fit Digital Thermometer Probe For Oral
Use

Cover To Fit Digital Thermometer Probe For Oral Use

2 11_0302_081 2300000 EACH

الشروط الخاصة التسلسل

- 1 - All equipment must be latest models of the range specified by the technical specifications (authorised certificate by the manufacturer must be submitted), old models will be rejected.
- 2 - Submission of a formally endorsed certificate of FDA approval/ clearance for selling equipment of USA origin to USA healthcare facilities (installation references must be included in the offered documents), CE mark (MDD) certification for European Equipment that meet the country of origin requirements clearly indicating the certifying body number for class 2 equipment. JIS (Japanese Industrial standards) approval for Japanese equipment, CSA (Canadian Standards Association) approval for Canadian equipment, offers not including such certificates will be rejected.
- 3 - Country of origin: USA, Canada, Japan, UK, Sweden, Finland, Denmark, Switzerland, Belgium, Germany, France, Netherlands.
- 4 - Products manufactured by well-known companies based in countries mentioned in term 3 and have manufacturing plants in different parts of the world and holds a free-sale certificates in the original country where the mother company will be taken into consideration.

الشروط الخاصة التسلسل

- 5 – For countries not mentioned in 3, ONE of the following conditions must be met: A formally endorsed certificate from the Institute of Biomedical Technology/ Royal Medical Services, indicating that the system itself or an older model of the same type has been used by the Royal Medical Services for a minimum period of five consecutive years with an acceptable service record. The offered model to have successfully passed through the process of medical equipment assessment programme in accordance with the regulations of the assessment committee in the Royal Medical Services (acceptance certificate must be submitted).
- 6 – A list of Reputable institutions worldwide where the same offered model is operating especially in the USA or Germany. Reference sites will be verified and evaluated at the discretion of the purchasing committee.
- 7 – Tender awards will be based upon the best value for money and best performance expectations (reputation, after sale support etc.).
- 8 – Offers must include a two-year full manufacturer warranty, including batteries labour and parts, if any.
- 9 – Clear technical brochures and data sheets must be delivered with each unit.
- 10 – Pricing must include services of sale, shipment, transportation, delivery from port to main medical stores or site ,installation and bringing into servces of the equipments
- 11 – Power requirements: single phase 220V, 50Hz.where applicable.

Probe Covers To Fit Digital Thermometer P45-2018-2

الشروط الخاصة للعطاء رقم

الشروط الخاصة التسلسل

- | | |
|----|---|
| 12 | - Offers must include clear technical brochures and data sheets. |
| 13 | - Offers must include compliance sheets in tabular format listing our required specification on one side, and a clear and unequivocal response to each point on the other side with reference to page and line numbers in the specific technical brochure(s). Offered items not complying with this point will be rejected immediately. |
| 14 | - Any accessories and consumable items necessary to operate the offered equipment(s) must be clearly identified and priced. Prices must be fixed for a minimal period of 5 years. |
| 15 | - Training: On site user and service training |
| 16 | - The possibility of further purchases of any quantity of the same item mentioned in the tender with the same terms and prices at any later time. |
| 17 | - In the event of any malfunction or defect at any time in any device purchased in the tender, a functioning replacement new device must be provided within a maximum time limit of three working days of the time of being informed of the defect. Failure to comply will impose a minimum penalty equal to the cost of consumables designated to each device. |
| 18 | - All devices are on free of charge basis and the company is responsible for the full maintenance during the tender period |
| 19 | - Quantity of Digital thermometers (200) on free of charge basis delivered to main medical stores of the Royal Medical Services. |

الشروط الخاصة التسلسل

- 20 - ALL GOODS ARE TO BE FRESHLY PREPARED UPON DISPATCH/ DELIVERY AND SHOULD HAVE AT LEAST 2/3 OF ITS SHELF LIFE. ANY ITEM CARRYING LESS THAN 2/3 OF ITS SHELF LIFE UPON RECEIPT SHOULD BE ACCOMPANIED WITH A CONFORMATION THAT YOU ACCEPT TO REPLACE ANY REMAINING UNUSED QUANTITIES AT YOUR EXPENSES AFTER THE EXPIRY DATE. SAID CONFIRMATION IS SUBJECT TO THE APPROVAL OF THE DIRECTOR OF THE ROYAL MEDICAL SERVICES AND WILL INCUR A FINE WHICH WILL BE DECIDED LATER BY DRMS ACCORDING TO THE LOSS THAT THIS DISCREPANCY WITH TERMS OF THE TENDER HAS CAUSED.
- 21 - Delivered consumables should show description of goods, manufacturing date, expiry date, batch no, storage conditions, stamped "Sold to RMS", and any other necessary information.
- 22 - Goods must be delivered under the same storage conditions that comply with their nature.
- 23 - All parcels, packages and related documents should be stamped or labeled with the following:-
DRMS/TENDER No. P45-2018-2
GHQ/JORDAN ARMED FORCES
AMMAN, JORDAN
- 24 - "SOLD TO DRMS" SHOULD BE PRINTED OR STAMPED ON ALL OUTER AND INNER INDIVIDUAL PACKS WITH INK NOT EASY TO REMOVE EITHER BY THE COMPANY OR BY THE LOCAL AGENT.

الشروط الخاصة التسلسل

25	- Probe covers to be delivered in two equal shipments (4-5) months apart, first shipment to be as soon as possible including the total quantity of the Digital thermometer.
26	- After arrival to Jordan, DRMS is not responsible for any handling charges for goods (like charges for issuing delivery note vouchers, unloading charges, local shipping agent's charges). GHQ/ RMS are not responsible for any demurrage charges and any other charges incurred by the port corporation caused by delay in presenting all necessary shipping document for clearing the goods.
27	- DRMS has the right to increase or decrease the awarded quantities by a percentage not exceeding 30% after final order notification with the same prices, terms and conditions of the contract upon request and contractors approval.
28	- Offers must include an original copy of free-sale certificates in the original country where the mother company is based endorsed by the ministry of health of that country.
29	- Spare parts availability must be guaranteed for a minimum of 10 years starting from date of installation and commissioning.

الشروط الخاصة التسلسل

- | | |
|----|---|
| 30 | - Offers must include a comprehensive, priced list of spare parts valid for a minimal period of 5 years. Any item not listed will be considered as free of charge. For the final list offers having equal chances of winning the award, the warding process will be based on the accumulative value of both the offered system and its running cost over the period of five years. Running cost include the approximate value of consumables, spare parts and accessories needed to operate the system over the period of five years. Only offers with the lowest sum value (system price and its running cost) will qualify for the award. Another copy of the list must be submitted at the Institute of Biomedical Technology (IBMT) before an installation-and-commissioning certificate can be issued. |
| 31 | - Bidders must submit their reservations / queries regarding tender specifications and / or special terms within the first third of the tender closing period. |
| 32 | - Goods must be previously evaluated and approved by DRMS or purchased by tenders. |



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item -1	Digital thermometer	Qty (3)
---------	---------------------	---------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N) , Notes	Brochure Page No.
1	Thermometers should be highly accurate and easy to use for measuring patient temperatures by oral, axillary and rectal means.		
2	To be used for adults, paediatrics, infants and neonatal.		
3	All components including probe covers must be homogeneous and ergonomically designed with no sharp edges.		
4	The display should read the temperature in Celsius and Fahrenheit.		
5	Light weight and easy to carry.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N) , Notes	Brochure Page No.
6	LCD display with backlight display.		
7	Thermometer probe and tips should be able to clean by using alcohol or soap and water.		
8	Water resistant.		
9	Measurements range: (32-43) C , (89.6-109.4)F.		
10	Operating modes: a. Quick mode b. Direct mode c. Cold mode		
11	Response time: a. Quick mode: less than 15 seconds b. Direct mode: less than 120 seconds c. Cold mode: less than 20 seconds		
12	Accuracy: a. Direct/cold mode: (+- 0.1C) b. Quick mode: (+- 0.3 C)		
13	Auto shut-off.		
14	Acoustic alarm signal when measurement is complete.		
15	All types of probes and covers should be quoted.		
16	Battery operated type.		
17	If the unit operated by rechargeable batteries, then the unit must be provided with charger.		
18	Battery charger indicator for rechargeable type.		
19	Low battery and dead battery indicator.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N) , Notes	Brochure Page No.
20	Long life battery : approx. 4000 takes or better.		
21	Two backup set of batteries must be included.		
22	Any other accessories should be quoted and priced separately.		
23	Samples must be submitted at DRMS for evaluation purposes. The purchasing committee shall have the right to reject unsatisfactory samples, and subsequently, the whole offer.		