

DIRECTORATE ROYAL MEDICAL SERVICES

مستهلكات قسطره الاشعه التداخليه

TENDER NO P100-2018-I-2

CORONARY ANGIOGRAPHY CATHETERS

Catheter, Angiographic (Polyethelene) Renal.

1	20_10_1361	360 EACH	Size 5.0 FG To Permit Guide Wire Size .(0.038"), Length (60-80cm), Fitting End M, Side Holes 0
---	------------	----------	--

VERTEBRAL CURVE CATHETER RADIO-OPAQUE TIP

2	20_10_1472	2400 EACH	size 5.0 FR length 80-100cm accep 0.038" G.WIRE
---	------------	-----------	--

CORONARY ANGIOGRAPHY CATHETERS

VERTEBRAL CURVE CATHETER ,Diagnostic catheter

3	20_10_1473	240 EACH	size 4.0 FR length 80-100cm braided one to one torque accep 0.038" Guide WIRE, WITH RADIO OPAQUE TIP
---	------------	----------	---

CARDIOVASCULAR CATHETERIZATION DISPOSABLES

CATHETER MULTIPURPOSE CURVE, OPEN ENDS WITH SIDE HOLES
ONE TO ONE TORQUE, BRAIDED with radio opaque tip, accep
0.038" WIRE.

4	20_10_148	400 EACH	5F Size length 65cm
---	-----------	----------	---------------------

CORONARY ANGIOGRAPHY CATHETERS

Catheters, Angiographic (Selective Catheters) Renal,
Sterile.

5	20_10_1671	36 EACH	4French, (> Or =15ml\Sec), Low Profile, Radio-Opaque Tip, 1:1 Torque, Permit G.Wire at least 0.038", No Side Holes, Length 60-80cm.
---	------------	---------	--

Catheters, Angiographic (Selective Catheters) Mani,
Sterile.

6	20_10_1673	60 EACH	5French, 1:1 Torque, Low Profile, Radio- Opaque tip, Permit G.Wire At Least 0.038", No Side Holes, Length (90-120cm)
---	------------	---------	--

5French, 1:1 Torque, Low Profile, Radio-Opaque, Tip
Marker, accep Guide Wire 0.038" Length 90-100cm

7	20_10_1674	100 EACH	Catheters, Angiographic (Selective Catheters) Head Hunter,
---	------------	----------	---

Catheters, Angiographic (Selective Catheters) Simmons
II &/ or III, Sterile.

8	20_10_1677	300 EACH	5French, Low profile, 1:1 Torque, Radio- Opaque tip, Permit G.Wire 0.038", No Side Holes, Length 90-100cm.
---	------------	----------	--

Catheters, Angiographic (Selective Catheters) Cobra

9	20_10_1678	60 EACH	5 French, Low profile, 1:1 Torque tip, Permit G.Wire 0.038", No Side Holes, Length 90-100cm.
---	------------	---------	--

DIRECTORATE ROYAL MEDICAL SERVICES

مستهلكات قسطره الاشعه التداخليه

TENDER NO P100-2018-I-2

Item No	Stock No	REQ QTY	UNIT	DESCRIPTION
CORONARY ANGIOGRAPHY CATHETERS				
				CATHETER, INTRODUCER SHEATH SET WITH HEMOSTATIC VALVE SIDE ARM INTRODUCING WIRE STERILE, SIZE 4Fr LENGTH (11- 13CM) ACCEPT G.WIRE 0.035"
10	20_10_179	300	EACH	
CARDIOVASCULAR CATHETERIZATION DISPOSABLES				
				CATHETER INTRODUCER SET EITHER ARTERIAL OR VENOUS ACCESS WITH HEMOSTASIS VALVE 12-16CM TEFLON SHEATH AND 17CM DILATOR FRENCH SIZE ADULT WITH STERILE ARM
11	20_10_181	2400	EACH	5F ACCEPT 0.035" WIRE
				CATHETER INTRODUCER SET EITHER ARTERIAL OR VENOUS ACCESS WITH HEMOSTASIS VALVE 10-13CM TEFLON SHEATH AND 17CM DILATOR FRENCH SIZE ADULT
12	20_10_182	1800	EACH	6F ACCEPT 0.035"-0.038" WIRE
CORONARY ANGIOGRAPHY CATHETERS				
				(GUIDE WIRE) HYDROPHILIC POLYMER -COATED WIRE
13	20_10_224	420	EACH	0.035" Size 180 length straight
				(GUIDE WIRE) HYDROPHILIC POLYMER -COATED WIRE
14	20_10_226	6000	EACH	0.035" Size 180 length Angled
CARDIOVASCULAR CATHETERIZATION DISPOSABLES				
				(GUIDE WIRE) HYDROPHILIC POLYMER -COATED WIRE
15	20_10_228	200	EACH	0.035" Size 260 length Angled
				0.014-0.035" Size 180-300CM length Angled or straight tip
16	20_10_2280	200	EACH	(GUIDE WIRE) HYDROPHILIC DISTAL 25CM STIFF WITH SOFT PROXIMAL PART

DIRECTORATE ROYAL MEDICAL SERVICES

مستهلكات قسطره الاشعه التداخليه

TENDER NO P100-2018-I-2

Item No	Stock No	Quantity	Description
			Hydrophilic wire Angled tip ,45-90 stiff, 180-200cm, with torque
17	20_10_2281	160 EACH	
			Hydrophilic guide wire straight tip , stiff care ,180-200 cm with torque
18	20_10_2282	200 EACH	
			Hydrophilic wire Angled stiff care, 260 cm all wire hydrophilic with torque
19	20_10_2283	1600 EACH	
			SUPER STIFF HYDROPHILIC COATED GIUDE WIRE TERUMO OR ITS EQUIVALENT
20	20_10_2284	600 EACH	0.035" 260cm length straight tip.
			SUPER STIFF HYDROPHILIC COATED GIUDE WIRE TERUMO OR ITS EQUIVALENT
21	20_10_2285	800 EACH	0.035" 260cm length Angled tip
			HYDROPHILIC COATED GIUDE WIRE TERUMO OR ITS EQUIVALENT
22	20_10_2288	120 EACH	0.035" 180-260cm straight distal shapable tip
CORONARY ANGIOGRAPHY CATHETERS			
			Hydrophilic guide wire STRAIGHT
23	20_10_322	60 EACH	shaft 0.035" with distal sigment tapers to 0.18, length 180cm, with torque device
			0.14 length 190cm
24	20_10_350	60 EACH	Embolic protection device (carotid filter wire) & free cell space (40-90) micron
			Guiding Catheters , Sterile, vertebral shape tipped, 6 Fr, Guiding, max. inner diameter > or = 0.070" with soft radio-opaque tip, length 120-140 cm.
25	20_10_3691	120 EACH	
CARDIOVASCULAR CATHETERIZATION DISPOSABLES			
			INTRODUCER SET DOUBLE DILATOR SYSTEM WITH SIDE ARM VALVE DILATOR LENGTH 31CM. AND INTRODUCER SHEATH LENGTH 23CM
26	20_10_497	96 EACH	6F french Size
27	20_10_498	48 EACH	7F french Size

DIRECTORATE ROYAL MEDICAL SERVICES

مستهلكات قسطره الاشعه التداخليه

TENDER NO P100-2018-I-2

Item No	Stock No	QTY	UNIT	DESCRIPTION
				Plug- In Y-Adaptor 2 ways, adapting proximal end of catheter , with side arm tube
28	20_10_5051	600	EACH	Y-Shape adaptor
				2 ways-adapting proximal end of the catheter without side arm
29	20_10_5052	300	EACH	Y-shape adapter
				Rotating Y-Adaptor 2 ways adapting proximal catheter or balloon
30	20_10_5053	48	EACH	Y-shape adapter
CARDIOVASCULAR CATHETERIZATION DISPOSABLES				
				PERCUTANEUOS TRANSLUMINAL CORONARY ANGIOPLASTY ACCESSORIES
31	20_10_508	1600	EACH	coronary angioplasty inflation device 20ml syringe up to 20 bar
CORONARY ANGIOGRAPHY CATHETERS				
				Micro mesh self expandable carotide stent, double layer, braided mesh nitinol, diameter 5-10mm, length 30-50mm ,Rapd exchange.
32	20_10_555	60	EACH	
				Catheter, Drainage All Purpose Set With Stiffening Cannula Trocar needle Stylet With Slippery Coat, Locking Mechanism & Pigtail, Sterile
33	20_10_813	480	EACH	Size 10 FG, trocar stylet with puncturing needle, hydrophilic coating, radio-opaque marker at drainage loop, self locking mechanism.
34	20_10_814	360	EACH	Size 12 FG, trocar stylet with puncturing needle, hydrophilic coating, radio-opaque marker at drainage loop, self locking mechanism.
Interventional Neuroradiology				
				5-6 Fr peripheral support catheter, double distal braided low profile, hydrophilic coated tapered tiper adapt 0.014- 0.035'' wire, 60-150 cm, with radio-opaque makers adaptable with 4 Fr sheah
35	20_10_83	600	EACH	

DIRECTORATE ROYAL MEDICAL SERVICES

مستهلكات قسطره الاشعه التداخليه

TENDER NO P100-2018-I-2

Item No	Quantity	Unit	Description
CORONARY ANGIOGRAPHY CATHETERS			
			5-6Fr, radio-opaque tip , 40-50 cm, adapt 0.038" Guide wire. with distal radio-opaque maker tip
36	20_10_863	1200 EACH	Manipulation catheter,
37	20_10_8651	4400 EACH	GENERAL PURPOSE DRAPE TO CONTAIN : 1- WATER RESISTANT ANGIOGRAPHY DRAPE, 200X300cm 2-FLOUROSCOPY COVER 90X90cm, 3-BACK TABLE COVER 100X200cm 4-WRAP 100X100cm 5-TABLE COVER 100X150cm 6-GALLIPOT 2X250ml / 1X2000ml (for guide wire) / 1X500ml 7- 2 reinforced gowns XL sterile Nylon cover for shield protection (65x85) cm
			Probe Ultra Sound Transducer Cover, Sterile
38	20_10_867	4800 EACH	.Size 14x147 CM
			Intracranial Micro Guide Wire Gold Tip, Sterile
39	20_10_880421	240 EACH	Double curve Hydrophilic intracranial guide wire, gold tip 0.014
Interventional Neuroradiology			
			Micro guide wire for neuroapplication, stainless steel, hydrophilic coated.
40	20_10_88071	360 EACH	Shape adaptable, 0.008, 1:1 torque, length 180-210 cm, radio-opaque, soft tip (30-40 cm).
Coils-Intracranial			
			size 1.5-20 mm loop diameter, 1-50cm length, variable helical loops , soft and waves
41	20_10_89022	50 Coils	Hypo-tube, intracranial platinumium or stainless steel coils with rapid detachment less than 3 seconds,
CORONARY ANGIOGRAPHY CATHETERS			
42	20_10_89113	60 EACH	Remodelling Intracranial balloon, Nylon, latex or silicon with two radio-opaque markers, semi-compliant, over 0.014 wire, 3-5 mm diameter, length 10-30 mm.
			Microspheres Well- Calibrated, 700-1300 microm.
43	20_10_9868	60 EACH	

الشروط الخاصة للعطاء رقم P100-2018-I-2 مستهلكات قسطره الاشعه التداخليه

الشروط الخاصة التسلسل

- 1 - All goods are to be freshly prepared on dispatch/delivered & should have at least 2/3 its shelf life upon receipt . Any item carrying less than 2/3 of its shelf life upon receipt should be accompanied with a confirmation that you accept to replace any remaining unused quantities at your expenses after the expiry date. Said confirmation is subject to the approval of the director of the Royal Medical Services and will incur a fine ,which will be decided later by DRMS according to the loss that this discrepancy with terms of the tender has caused.
- 2 - Description of goods ,expiry date ,lot number as well as manufacturing date must be mentioned clearly on the inner & outer packs.
- 3 - RMS has the right to exchange any size of the awarded items with any other size such as exchange should be effected within(24 hours) after a written request from RMS
- 4 - All goods should be FDA or CE approved
- 5 - Goods should be delivered under the same storage conditions that comply with their nature,storage conditions must be mentioned clearly on all documents.
- 6 - RMS has the right to increase the awarded items by 30% with the same prices and terms.and conditions.
- 7 - Pricing must include services of sale, shipment, transportation, and delivery From. port to. main medical store.
- 8 - Quoted items should be previously approved by DRMS.

الشروط الخاصة للعتاء رقم P100-2018-1-2 مستهلكات قسطره الاشعه التداخليه

الشروط الخاصة التسلسل

- 9 - Delivery :In 2 Consignments (4-5) months apart ,allowing partial delivery first consignment to be within one month from final order
- 10 - Samples of the items must be submitted along with the offers, showing the same serial No./item No. as per our tender documents specifications.
- 11 - Custom clearance of goods shall be the responsibility of the Jordanian armed forces(JAF)however, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the ports corporation (including expenses caused by delay in presenting the necessary documents for either clearing or transporting the goods to the required location mentioned in the final order ,delivery note issuing charges, unloading charges, local shipping charges etc).the supplier is also responsible for providing of all relevant shipping documents ,together with the delivery order(s).
- 12 - After arrival to jordan DRMS is not responsible for any handling charges for goods(like charges for issuing delivery order for incoming shipment unloading charges,local shipping agents charges
- 13 - All parcels,packages&any relative documents to me marked:
D.R.M.S Tender NO P100-2018-1-2 G.H.Q Jordan Armed Forces
Amman-Jordan For Jordanian Armed Forces

بسم الله الرحمن الرحيم

صفحة رقم 3 - 3

القيادة العامة للقوات المسلحة الاردنية

مديرية الخدمات الطبية الملكية

التزويد

الشروط الخاصة للعطاء رقم P100-2018-1-2 مستهلكات قسطره الاشعه التداخليه

الشروط الخاصة التسلسل

14 - In case of any worldwide problem or recall of any awarded item. the company is obliged to replace any remaining quantities with equivalent or better materials otherwise DRMS has the right to buy these materials from another sources and impose you the differences of prices and any penalties according to the rules and regulations of DRMS

15 - If new generation appears after the final order, it must be exchanged with the old generation without affecting the prices within (4 weeks) from receiving the written request from KHMC Supply department